



INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30
Hillsborough, NJ 08844 USA
Phone: (908) 359-6778; Fax: (908) 359-1179

ANALYTICAL CONDITIONS

G C ANALYSIS

APIOLE

025390S Batch: 02

COLUMN	Carbowax Capillary 30 m x 0.32 mm x 0.25 µm
TEMPERATURE	Column : 70 to 220°C (5°C/min) Injector: 200°C Detector : 230°C
DETECTOR	FID
CARRIER GAS	Hydrogen 6psi
SAMPLE SOLVENT	Isopropanol

