



INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30
Hillsborough, NJ 08844 USA
Phone: (908) 359-6778; Fax: (908) 359-1179

G C ANALYSIS

alpha-CEDRENE

025432S Batch: 08

ANALYTICAL CONDITIONS

COLUMN **Carbowax Capillary**
 30 m x 0.32 mm x 0.25 µm

TEMPERATURE Column : 70 to 220°C (5°C/min)
 Injector: 200°C
 Detector: 230°C

DETECTOR FID

CARRIER GAS Hydrogen 6psi

SAMPLE SOLVENT Isopropanol

