



# INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: [www.indofinechemical.com](http://www.indofinechemical.com)

e-mail: [chemical@indofinechemical.com](mailto:chemical@indofinechemical.com)

05-2300 TRICOSANE

LOT NO: 2294:8

PURITY: 99 + %

GC    Injector : On column  
      Detector : FID  
      Column : DB-wax, 30m; 0.32mm  
      Carrier : He

### Temperature

#### Column

Initial: 130<sup>o</sup>C

Final : 240<sup>o</sup>C in 60 min

Rate : 3<sup>o</sup>C/min

Injector : 130<sup>o</sup>C

Detector : 250<sup>o</sup>C

