



INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: www.indofinechemical.com

e-mail: chemical@indofinechemical.com

05-2700 HEPTACOSANE

LOT NO: 6113-5

PURITY: 99 + %

GC Injector : On column
 Detector : FID
 Column : DB-wax, 30m; 0.32mm
 Carrier : He

Temperature

Column

Initial: 110 °C

Final : 220 °C

Rate : 3 °C/min

Injector : 80 °C + 140 °C/min Final 250 °C

Detector : 250 °C

