



# INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: [www.indofinechemical.com](http://www.indofinechemical.com)

e-mail: [chemical@indofinechemical.com](mailto:chemical@indofinechemical.com)

10-0900

NONANOIC ACID

LOT NO: 9171/6

PURITY: > 99%

GC    Injector : On column  
      Detector : FID  
      Column : DB-wax, 30m; 0.32mm  
      Carrier : He

#### Temperature

##### Column

Initial: 130°C

Final : 220°C in 60 min

Rate : 3°C/min

Injector : 130°C

Detector : 250°C

