



INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: www.indofinechemical.com

e-mail: chemical@indofinechemical.com

10-1811 Petroselinic Acid (6c)

GC Methylated sample

Lot no: O 321:11

Purity: 99 %

Storage: Freezer

GC: Injection : On column
Detector : FID
Column : DB-Wax 30m x 0.32mm
film 0.25um
Carrier : Helium 14 psi

Temperature

Column

Initial : 160°C 3 min

Final : 240°C

Rate : 3°C/min.

Injector : 240°C

Detector : 270°C

