



INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: www.indofinechemical.com

e-mail: chemical@indofinechemical.com

10-2005 Eicosapentaenoic Acid
(5,8,11,14,17-all cis)

Lot no: Q 026:6

Purity: 99+%

GC: Injection : On column
Detector : FID
Column : DB-Wax 30m x 0.32mm film 0.25um
Carrier : Helium 14psi

Temperature Column:

Initial : 190 C , 3min

Final : 240 C

Rate : 3 C /min

Injector : 240 C

Detector : 270 C

