



INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: www.indofinechemical.com

e-mail: chemical@indofinechemical.com

11-1600 Phytanic Acid

GC Methylated sample

Lot no: K 318:10

Purity: 97+ %

GC: Injection : On column
Detector : FID
Column : DB-Wax 30m x 0.32mm
 : film 0.25um
Carrier : Helium 14 psi

Temperature
Column
 Initial : 160°C 3 min
 Final : 240°C
 Rate : 3°C/min.
Injector : 240°C
Detector : 270°C

