



121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: [www.indofinechemical.com](http://www.indofinechemical.com)

e-mail: [chemical@indofinechemical.com](mailto:chemical@indofinechemical.com)

Lot no: F 246:6  
Purity: 98+ %

14-0903 3-Hydroxy-nonanoic Acid  
GC Methylated sample

GC: Injection : On column

Detector : FID

Column : DB-Wax 30m x 0.32mm

film 0.25um

Carrier : Helium , 1.6 bar

Temperature

Column

Initial :

80°C 3 min

Final :

240°C

Rate :

3°C/min.

Injector

:

80°C + 140°C/min.

Final 240°C

Final 240°C

0.0  
5.0  
10.0  
15.0  
20.0  
25.0  
30.0  
35.0  
40.0  
45.0