



INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: www.indofinechemical.com

e-mail: chemical@indofinechemical.com

14-1003 3-Hydroxydecanoic Acid

GC Methylated sample

Lot no: I 199:5

Purity: 98+ %

GC: Injection : On column
Detector : FID
Column : HP-5 30m x 0.32mm
film 0.25um
Carrier : Helium 14 psi

Temperature

Column

Initial : 80°C 3 min

Final : 260°C

Rate : 10°C/min.

Injector : 240°C

Detector : 270°C

