



INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: www.indofinechemical.com

e-mail: chemical@indofinechemical.com

14-1403 3-Hydroxytetradecanoic Acid

Lot no: E 305:9

GC Methylated sample

Purity: 99+ %

GC: Injection : On column
Detector : FID
Column : DB-Wax 30m x 0.32mm
film 0.25um
Carrier : Helium , 3.2 bar

Temperature
Column
Initial : 130°C 3 min
Final : 240°C
Rate : 3°C/min.

Injector : 130°C + 140°C/min. Final 240°C
Detector : 240°C

