



# INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: [www.indofinechemical.com](http://www.indofinechemical.com)

e-mail: [chemical@indofinechemical.com](mailto:chemical@indofinechemical.com)

15-0400 SUCCINIC ACID

LOT NO: 9524-12

GC Methylated Succinic acid

PURITY: >99%

GC    Injector    : On column  
      Detector    : FID  
      Column     : DB-wax, 30m; 0.32mm  
      Carrier    : He

#### Temperature

##### Column

Initial : 34<sup>o</sup>C 3 min

Final    : 220<sup>o</sup>C

Rate     : 3<sup>o</sup>C / min

Injector : 130<sup>o</sup>C

Detector : 250<sup>o</sup>C

