



INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: www.indofinechemical.com

e-mail: chemical@indofinechemical.com

20-0500 Methyl Pentanoate

Lot no: B 462:12

Purity: 99+ %

GC: Injection : On column
Detector : FID
Column : DB-Wax 30m x 0.32mm
 film 0.25um
Carrier : Helium , 1.6 bar

Temperature

Column

Initial : 34°C 3 min

Final : 240°C

Rate : 3°C/min.

Injector : 40°C + 140°C/min. Final 240°C

Detector : 240°C

