



# INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: [www.indofinechemical.com](http://www.indofinechemical.com)

e-mail: [chemical@indofinechemical.com](mailto:chemical@indofinechemical.com)

20-2300 Methyl Tricosanoate

Lot no: N 009:1

Purity: 99+ %

**GC:** Injection : On column  
Detector : FID  
Column : HP-5 30m x 0.32mm  
film 0.25um  
Carrier : Helium 14 psi

Temperature  
Column

Initial : 120°C 3 min

Final : 260°C

Rate : 10°C/min.

Injector : 240°C

Detector : 270°C

