



INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: www.indofinechemical.com

e-mail: chemical@indofinechemical.com

Certificate of Analysis

Product no:	20-2300
Product Name:	Methyl Tricosanoate
Lot no:	Q 208:11
Purity:	>99%
Analysis:	GC
MW:	368,7
Storage:	Freezer



INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: www.indofinechemical.com

e-mail: chemical@indofinechemical.com

Certificate of Analysis

Product	20-2300	Methyl Tricosanoate
Lot	Q 208:11	
Mw	368.6	
Purity	99% by GC	
Supplied as	Neat	
Stability	min 5 years in refrigerator	

GC:	Injection: On column
	Detector: FID
	Column: HP-5, 15m x 0,32mm film 0,25um
	Carrier: Helium 14 psi

Temperature column:
Initial: 160C, 3min
Final: 260C
Rate: 7C/min
Injector: 240C
Detector: 270C

