



# INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: [www.indofinechemical.com](http://www.indofinechemical.com)

e-mail: [chemical@indofinechemical.com](mailto:chemical@indofinechemical.com)

21-2019 METHYL 19-METHYLEICOSANOATE

LOT NO: 9752-7

PURITY: > 98 %

GC    Injector : On column  
      Detector : FID  
      Column  : DB-wax, 30m; 0.32mm  
      Carrier  : He

#### Temperature

##### Column

Initial: 130<sup>o</sup>C

Final  : 220<sup>o</sup>C in 60 min

Rate   : 3<sup>o</sup>C/min

Injector : 130<sup>o</sup>C

Detector  : 250<sup>o</sup>C

