



INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: www.indofinechemical.com

e-mail: chemical@indofinechemical.com

25-1000 DIMETHYLSEBACATE

LOT NO: A126/5

PURITY: +98%

GC Injector : On column
 Detector : FID
 Column : DB-wax, 30m; 0.32mm
 Carrier : He

Temperature

Column

Initial: 130^oC

Final : 220^oC

Rate : 1^oC/min

Injector : 130^oC

Detector : 250^oC

