



# INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: [www.indofinechemical.com](http://www.indofinechemical.com)

e-mail: [chemical@indofinechemical.com](mailto:chemical@indofinechemical.com)

31-1001 1-MONODECANOIN

LOT NO: 7082:3

GC Silylated with BSTFA

PURITY: 99 %

GC Injection : On column  
Detector : FID  
Column : CP-WAX-52 CB, 30m x 0.32mm  
film 0.25um  
Carrier : Helium

#### Temperature

##### Column

Initial : 110°C

Final : 240°C

Rate : 3°C/min

Injector : 110°C + 140°C/min. Final 240°C

Detector : 240°C

