



# INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: [www.indofinechemical.com](http://www.indofinechemical.com)

e-mail: [chemical@indofinechemical.com](mailto:chemical@indofinechemical.com)

40-1800 OCTADECANOL

LOT NO: 9124:4

PURITY: 99+%

GC Injection : On column  
Detector : FID  
Column : DB - Wax 30m x 0.32mm  
film 0.25um  
Carrier : Helium 1.6 bar

#### Temperature

##### Column

Initial : 150°C 3 min

Final : 240°C

Rate : 3°C/min

Injector : 150°C + 140°C/min. Final 240°C

Detector : 240°C

