



INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: www.indofinechemical.com

e-mail: chemical@indofinechemical.com

40-2000 Eicosanol

Lot no: N 155:5

Purity: 99 %

GC: Injection : On column
Detector : FID
Column : HP-5 30m x 0.32mm
film 0.25um
Carrier : Helium 14 psi

Temperature

Column
Initial : 120°C 3 min
Final : 260°C
Rate : 10°C/min.
Injector : 240°C
Detector : 270°C

