



INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: www.indofinechemical.com

e-mail: chemical@indofinechemical.com


40-2000 Eicosanol

Lot no: Q118:8

Purity: 99+%

GC: Injection : On column
Detector : FID
Column : HP-5 15m x 0.32mm film 0.25um
Carrier : Helium 14psi

Temperature Column:
Initial : 160 C , 3min
Final : 260 C
Rate : 7 C /min
Injector : 240 C
Detector : 270 C



#	Time	Area	Height	Width	Area%	Symmetry
1	3.393	22.7	16.3	0.0232	0.856	0.716
2	5.309	2623.5	609.6	0.0662	99.144	1.112