



INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: www.indofinechemical.com

e-mail: chemical@indofinechemical.com

90-1052 Mixture BR 2

Lot no: I 169:4

GC Injection : On column
Detector : FID
Column : DB-Wax 30m x 0.32m
film 0.25um
Carrier : Helium, 14 psi

Temperature

Column
Initial : 80°C 3 min.
Final : 240°C
Rate : 5°C / min.
Injector : 240°C
Detector : 270°C

Peak	Compound	Weight%
1	Methyl 10-methyldodecanoate	12.5
2	Methyl tridecanoate	12.5
3	Methyl 12-methyltridecanoate	12.5
4	Methyl tetradecanoate	12.5
5	Methyl 12-methyltetradecanoate	12.5
6	Methyl pentadecanoate	12.5
7	Methyl 14-methylpentadecanoate	12.5
8	Methyl hexadecanoate	12.5

100.0

