



# INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: [www.indofinechemical.com](http://www.indofinechemical.com)

e-mail: [chemical@indofinechemical.com](mailto:chemical@indofinechemical.com)

90-1053 Mixture BR 3

Lot no: N 005:1

GC

Injection : On column  
Detector : FID  
Column : DB-Wax 30m x 0.32m  
film 0.25um  
Carrier : Helium, 14 psi

Temperature

Column

Initial : 130°C 3 min.

Final : 240°C

Rate : 3°C / min.

Injector : 240°C

Detector : 270°C

Peak	Compound	Weight%
1	Methyl 12-methyltetradecanoate	12.5
2	Methyl pentadecanoate	12.5
3	Methyl 14-methylpentadecanoate	12.5
4	Methyl hexadecanoate	12.5
5	Methyl 14-methylhexadecanoate	12.5
6	Methyl heptadecanoate	12.5
7	Methyl 16-methylheptadecanoate	12.5
8	Methyl octadecanoate	12.5

100.0

