



INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: www.indofinechemical.com

e-mail: chemical@indofinechemical.com

90-1053 Mixture BR 3

Lot no: N 005:1

GC	Injection : On column	Peak	Compound	Weight%
	Detector : FID	1	Methyl 12-methyltetradecanoate	12.5
	Column : DB-Wax 30m x 0.32m film 0.25um	2	Methyl pentadecanoate	12.5
	Carrier : Helium, 14 psi	3	Methyl 14-methylpentadecanoate	12.5
	Temperature	4	Methyl hexadecanoate	12.5
	Column	5	Methyl 14-methylhexadecanoate	12.5
	Initial : 130°C 3 min.	6	Methyl heptadecanoate	12.5
	Final : 240°C	7	Methyl 16-methylheptadecanoate	12.5
	Rate : 3°C / min.	8	Methyl octadecanoate	12.5
	Injector : 240°C			100.0
	Detector : 270°C			

