



INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: www.indofinechemical.com

e-mail: chemical@indofinechemical.com

90-1054 Mixture BR 4

GC Injection : On column
Detector : FID
Column : DB-Wax 30m x 0.32m
film 0.25um
Carrier : Helium, 14 psi

Temperature

Column

Initial : 130°C 3 min.

Final : 240°C

Rate : 3°C / min.

Injector : 240°C

Detector : 270°C

Lot no: I 168:4

Peak	Compound	Weight%
1	Methyl 14-methylhexadecanoate	12.5
2	Methyl heptadecanoate	12.5
3	Methyl 16-methylheptadecanoate	12.5
4	Methyl octadecanoate	12.5
5	Methyl 16-methyloctadecanoate	12.5
6	Methyl nonadecanoate	12.5
7	Methyl 18-methylnonadecanoate	12.5
8	Methyl eicosanoate	12.5

100.0

FID1 A, (REF_MIXLOTNO_I_168_400.D)

