



INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: www.indofinechemical.com

e-mail: chemical@indofinechemical.com

05-2000 EICOSANE

LOT NO: 4203:5

PURITY: 99 + %

GC Injection : On column
Detector : FID
Column : CP-WAX-52 CB, 30m x 0.32mm
film 0.25um
Carrier : Helium

Temperature

Column

Initial : 110°C

Final : 240°C

Rate : 3°C/min

Injector : 110°C + 140°C/min. Final 240°C

Detector : 240°C

