



INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: www.indofinechemical.com

e-mail: chemical@indofinechemical.com

05-2400 TETRACOSANE

LOT NO: 9203-5

PURITY: >99%

GC
Injector : On column
Detector : FID
Column : DB-wax, 30m; 0.32mm
Carrier : He

Temperature

Column

Initial: 130°C

Final : 220°C in 60 min

Rate : 3°C/min

Injector : 130°C

Detector : 250°C

